EPC 136-74-11

HILLSBOROUGH COUNTY
ENVIRONMENTAL PROTECTION COMMISSION
906 JACKSON STREET
TAMPA, FLORIDA 33602

WATER QUALITY REPORT

Sample location Bonnie Creek & N. 301 Date 7-22-76

Collector (s) L. Baker, J. Myers Time 1245

Dated Received 7-22-76 By Linda Powell

Remarks

Sample #	Parameter	Value
1	MENDO ± /100 ml	16000
	MF-C ± /100 ml.	1000
2	~~M-ENDO ± /100 ml.~~	
	~~MF-C ± /100 ml.~~	

Analyst Linda Powell
Date 7-23-76

EPC-169-76-11

HILLSBOROUGH COUNTY ENVIRONMENTAL PROTECTION COMMISSION
BACTERIOLOGICAL RESULTS
MEMBRANE FILTER DATA SHEET

Location 6 MILE CREEK & N. 301 Coll. Date 7-22-76 Time 1045
Collector(s) R. BOLER Analyzed by LP Date 7-22
Received by L. POWELL Time 1245 Date 7-22

Code: TNTC Too Numerous to Count
T Typical Colonies
AT Atypical Colonies
NC No Colonies

Chlorine _____ mg/l
S_2O_3 in Bottle? _____
Incub. Time: In [ ] Out [ ]

MF NO.	MEDIUM	SAMPLE VOLUME ML.	COUNT	REMARKS
1	M-ENDO	.01	16	TAT
2	"	.01	1	TAT
1	MF-C	0.1	1	TAT
2	"	.01	1	TAT

35.5°C Final Presumptive Coliform Count 16000 Colonies/100 ml.

Calculation: $\frac{\text{No. of Colonies of Indicator Organism}}{\text{No. of ml. of Sample Filtered}} \times 100 = \text{No. of Colonies/100 ml. Sample}$

44.5°C Final Confirmed Fecal Coliform Count 1000 Colonies/100 ml.