Hillsborough County Environmental Protection Commission
Bacteriological Results Form B-2
Membrane Filter Data Sheet

Location Six-Mile Creek Sta. #4 Date 8/22/24 Time 1030
Collector(s) Kavouras/Lewis Analyzed by JC Date 8/22/24
Received by Janene Caldwell Time 1210 Date 8/22/24
Code: TNTC Too Numerous to Count
T Typical colonies
AT Atypical colonies
NC No colonies
Chlorine _____ mg/l
S_2O_3 in bottle? yes

MF No.	Medium	Sample volume ml.	Count		Remarks
1	m-Endo	1.0	17		TAT
2	2	0.1	7		TAT
3		0.01	0		AT
1	MF-C	1.0	1		TAT
2	2	0.1	0		NC
3		0.01	0		NC
					Filters Sterilized too long

24 hour Final Presumptive Coliform Count 1700 No. of colonies per 100ml.
Calculation: $\frac{\text{No of colonies of indicator organism (m-Endo)}}{\text{No of ml of Sample filtered}} \times 100 =$ No. of colonies in 100ml sample

24 hour Final Confirmed Fecal Coliform Count 100 No. of colonies per 100ml